



CERTIFICATE OF ANALYSIS

**Product Description:** Calibrator 11  
**Catalog Number:** C0244-2  
**Lot Number:** A00U0424  
**Expiration Date:** 28 FEB 2030  
**Storage Temperature:** 2-8°C  
**Intended Assay Species:** Human, Mouse, NHP

**Concentration Assignment:**

Analyte	Amount of Analyte in Calibrator	Unit of Measure	Stock Concentration of Reconstituted Calibrator	Unit of Measure
TGF-β1	44125	pg	176500	pg/mL
TGF-β2	53750	pg	215000	pg/mL
TGF-β3	53625	pg	214500	pg/mL

**Reconstitution Volume:** 250 µL

To calculate the **Calibrator Standard 1 (top of curve) concentration**, please refer to the dilution-factor used in the **Prepare Calibrator Standards** section of the respective product insert. Divide the stock concentration of the reconstituted calibrator by this dilution factor to generate the working calibrator standard 1 concentration.

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
<b>Review/Approval</b>	Karen Hamilla	Quality		07 APR 2025