



### Extension Memo

**Product Description:** Neurodegeneration Control Pack 1 (4G8)  
**Lot Number:** CP000089  
**New Expiration Date:** 31 MAR 2027  
**Catalog Number(s):** C40RQ-1

Based on additional testing, the original date of expiration for this control pack has been extended by 60 months.

The expiration dates for the following components have also been extended:

Description	Lot Number	Original Expiration Date	New Expiration Date
Neurodegeneration Control 1 (4G8)	A00C0643	31 MAR 2022	31 MAR 2027
Neurodegeneration Control 2 (4G8)	A00C0644	31 MAR 2022	31 MAR 2027
Neurodegeneration Control 3 (4G8)	A00C0645	31 MAR 2022	31 MAR 2027

Function	Name	Signature	Date
Quality	Karen Hamilla		05 Feb 2026



## CERTIFICATE OF ANALYSIS

Page 1 of 1

**Control Pack Description:** Neurodegeneration Control Pack 1 (4G8)

**Catalog Number:** C40RQ-1

**Lot Number:** CP000089

**Expiration Date:** 31 MAR 2022

**Concentration Assignment: The following concentrations were measured using the A $\beta$  Peptide Panel 1 (4G8) V-PLEX<sup>®</sup> kit:**

Analyte	Control 1	Control 2	Control 3	Unit of Measure
A $\beta$ 40	7790	2650	417	pg/mL
A $\beta$ 38	3840	961	223	pg/mL
A $\beta$ 42	1490	516	205	pg/mL

**The following concentrations were measured using the Human Total Tau V-PLEX<sup>®</sup> kit:**

Analyte	Control 1	Control 2	Control 3	Unit of Measure
Tau	1330	415	161	pg/mL

### Components:

Description	Lot Number	Storage Temperature	Expiration Date
Neurodegeneration Control 1 (4G8)	A00C0643	$\leq -70^{\circ}\text{C}$	31 MAR 2022
Neurodegeneration Control 2 (4G8)	A00C0644	$\leq -70^{\circ}\text{C}$	31 MAR 2022
Neurodegeneration Control 3 (4G8)	A00C0645	$\leq -70^{\circ}\text{C}$	31 MAR 2022

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Review/Approval	Karen Hamilla	Quality		05 MAR 2024