

CERTIFICATE OF ANALYSIS

Page 1 of 1

Product Description: V-PLEX® and V-PLEX Plus Mouse IP-10 Kit
Kit Lot Number: K0082350
Expiration Date: 30 Sep 2025
Kit Catalog Numbers: K152NV-Series

Lot-Specific Components:

Description	Lot Number	Storage Temperature	Expiration Date
Mouse IP-10 SECTOR Plate	Z0047899	2-8°C	28 Feb 2026
Anti-ms IP-10 Antibody	D0082136	2-8°C	31 Jan 2027
Cytokine Panel 1 (mouse) Kit Calibrator	A0080353	2-8°C	29 Feb 2028
Diluent 41	M0410052	≤ -10°C	30 Sep 2025
Diluent 45	M0450052	≤ -10°C	30 Sep 2025

Additional components may be provided to execute the assay. Additional components are not lot-specific.

Precision and Accuracy Specifications:

Spot	Name	Calibrator		Cytokine Panel 1 (mouse) Control 1		Cytokine Panel 1 (mouse) Control 2		Cytokine Panel 1 (mouse) Control 3	
		Recovery	CV	Recovery	CV	Recovery	CV	Recovery	CV
1	IP-10	80% - 120%	< 20%	75% - 125%	< 20%	75% - 125%	< 20%	75% - 125%	< 20%


Spot	Name	LLOQ			ULOQ		
		Value	Recovery	CV	Value	Recovery	CV
1	IP-10	2.15	75% - 125%	< 20%	650	75% - 125%	< 20%

Stock Concentration of Calibrator:

Spot	Name	Conc.	Units
1	IP-10	1010	pg/mL

Dilution to Top of Curve = 1X

The above product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Review/Approval	Karen Hamilla	Quality		09 Nov 2023