

**CERTIFICATE OF ANALYSIS**

**Product Description:** V-PLEX® and V-PLEX Plus Vascular Injury Panel 2 (human) Kit  
**Lot Number:** K0082489  
**Expiration Date:** 31 DEC 2025  
**Catalog Number:** K15198-Series; K151SS-Series; K151SR-Series; K151SU-Series

**Lot-Specific Components:**

Description	Lot Number	Storage Temperature	Expiration Date
Anti-hu CRP Antibody	D0082276	2-8°C	30 NOV 2027
Anti-hu ICAM-1 Antibody	D0082278	2-8°C	30 NOV 2027
Anti-hu SAA Antibody	D0082275	2-8°C	30 NOV 2027
Anti-hu VCAM-1 Antibody	D0082277	2-8°C	30 NOV 2027
Vascular Injury Panel 2 (human) Plate	Z0057474	2-8°C	30 NOV 2026
Diluent 101	M1010052	≤ -10°C	31 MAY 2026
Vascular Injury Panel 2 (human) Calibrator Blend	A0080367	≤ -70°C	31 DEC 2025

Additional components may be provided to execute the assay. Additional components are not lot-specific.

**Precision, Accuracy, and Sensitivity Specifications:**

Spot	Name	Calibrator		Vascular Injury Control 1		Vascular Injury Control 2		Vascular Injury Control 3	
		Recovery	CV	Recovery	CV	Recovery	CV	Recovery	CV
1	SAA	80% - 120%	< 20%	75% - 125%	< 20%	75% - 125%	< 20%	75% - 125%	< 20%
2	CRP	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
3	VCAM-1	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
4	ICAM-1	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%

Spot	Name	LLOQ			ULOQ		
		Value	Recovery	CV	Value	Recovery	CV
1	SAA	54	80% - 120%	< 20%	138000	80% - 120%	< 20%
2	CRP	27.6	80% - 120%	< 20%	49600	80% - 120%	< 20%
3	VCAM-1	37.6	80% - 120%	< 20%	32000	80% - 120%	< 20%
4	ICAM-1	15	80% - 120%	< 20%	32700	80% - 120%	< 20%

**Non-Specific Binding Specifications:**

Spot	Name	Spec
1	SAA	< 1.0%
2	CRP	< 1.0%
3	VCAM-1	< 1.0%
4	ICAM-1	< 1.0%

**Stock Concentration of Calibrator:**

Spot	Name	Conc.	Units
1	SAA	4840000	pg/mL
2	CRP	3920000	pg/mL
3	VCAM-1	990000	pg/mL
4	ICAM-1	1006000	pg/mL

Dilution to top of Curve: 20X

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
<b>Review/Approval</b>	Karen Hamilla	Quality		22 JUL 2024