



### Extension Memo

Page 1 of 1

**Product Description:** V-PLEX® and V-PLEX Plus Human IL-10 Kit  
**Catalog Number(s):** K151QU-Series; K156QU-Series  
**Lot Number:** K0082525  
**New Expiration Date:** 28 Feb 2026

Based on additional testing, the original date of expiration for this kit has been extended by 12 months.

The expiration dates for the following components have also been extended:

Description	Lot Number	Original Expiration Date	New Expiration Date
Human IL-10 Plate	Z0047789	30 Apr 2025	28 Feb 2026
Diluent 2	M0020123	28 Feb 2025	28 Feb 2026
Diluent 3	M0030118	31 Mar 2025	28 Feb 2026

Function	Name	Signature	Date
Quality	Karen Hamilla		14 Feb 2025

**CERTIFICATE OF ANALYSIS**

**Product Description:** V-PLEX® and V-PLEX Plus Human IL-10Kit  
**Lot Number:** K0082525  
**Expiration Date:** 28 FEB 2025  
**Catalog Number:** K151QU-Series; K156QU-Series

**Lot-Specific Components:**

Description	Lot Number	Storage Temperature	Expiration Date
Human IL-10 Plate	Z0047789	2-8°C	30 APR 2025
Proinflammatory Panel 1(human) Calibrator Blend	A0080339	2-8°C	31 JUL 2027
Anti-hu IL-10 Antibody	D0081975	2-8°C	30 JUN 2026
Diluent 2	M0020123	≤ -10°C	28 FEB 2025
Diluent 3	M0030118	≤ -10°C	31 MAR 2025

Additional components may be provided to execute the assay. Additional components are not lot-specific.

**Precision, Accuracy, and Sensitivity Specifications:**

Spot	Name	Calibrator		Proinflammatory Panel 1 (human) Control 1		Proinflammatory Panel 1 (human) Control 2		Proinflammatory Panel 1 (human) Control 3	
		Recovery	CV	Recovery	CV	Recovery	CV	Recovery	CV
1	IL-10	80% - 120%	< 20%	90% - 130%	< 20%	90% - 130%	< 20%	90% - 130%	< 20%

Spot	Name	LLOQ			ULOQ		
		Value	Recovery	CV	Value	Recovery	CV
1	IL-10	0.298	80% - 120%	< 20%	233	80% - 120%	< 20%

**Stock Concentration of Calibrator:**

Spot	Name	Conc.	Units
1	IL-10	386	pg/mL

Dilution to top of Curve: 1X

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
<b>Review/Approval</b>	Karen Hamilla	Quality		24 SEP 2024