

**CERTIFICATE OF ANALYSIS**

**Product Description:** V-PLEX® and V-PLEX Plus Human MCP-1 Kit  
**Lot Number:** K0082533  
**Expiration Date:** 30 APR 2026  
**Catalog Number:** K151NN-Series; K156NN-Series

**Lot-Specific Components:**

Description	Lot Number	Storage Temperature	Expiration Date
Anti-hu MCP-1 Antibody	D0081963	2-8°C	31 MAY 2026
Chemokine Panel 1 (human) Calibrator Blend	A0080388	2-8°C	31 MAR 2029
Human MCP-1 Plate	Z0048009	2-8°C	31 JAN 2027
Diluent 3	M0030128	≤ -10°C	31 OCT 2026
Diluent 43	M0430055	≤ -10°C	30 APR 2026

Additional components may be provided to execute the assay. Additional components are not lot-specific.

**Precision, Accuracy, and Sensitivity Specifications:**

Spot	Name	Calibrator		Chemokine Panel 1 (human) Control 1		Chemokine Panel 1 (human) Control 2		Chemokine Panel 1 (human) Control 3	
		Recovery	CV	Recovery	CV	Recovery	CV	Recovery	CV
1	MCP-1	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%

Spot	Name	LLOQ			ULOQ		
		Value	Recovery	CV	Value	Recovery	CV
1	MCP-1	1.09	80% - 120%	< 20%	375	80% - 120%	< 20%

**Stock Concentration of Calibrator:**

Spot	Name	Conc.	Units
1	MCP-1	553	pg/mL

Dilution to top of Curve: 1X

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
<b>Review/Approval</b>	Karen Hamilla	Quality		15 OCT 2024