

**CERTIFICATE OF ANALYSIS**

**Product Description:** V-PLEX® and V-PLEX Plus Human VEGF Kit  
**Lot Number:** K0082557  
**Expiration Date:** 30 SEP 2026  
**Catalog Number:** K151RH-Series; K156RH-Series

**Lot-Specific Components:**

Description	Lot Number	Storage Temperature	Expiration Date
Anti-hu VEGF Antibody	D0082185	2-8°C	30 JUN 2027
Cytokine Panel 1 (human) Calibrator Blend	A0080383	2-8°C	31 DEC 2028
Human VEGF Plate	Z0057506	2-8°C	28 FEB 2027
Diluent 3	M0030129	≤ -10°C	30 NOV 2026
Diluent 43	M0430056	≤ -10°C	30 SEP 2026

Additional components may be provided to execute the assay. Additional components are not lot-specific.

**Precision, Accuracy, and Sensitivity Specifications:**

Spot	Name	Calibrator		Cytokine Panel 1 (human) Control 1		Cytokine Panel 1 (human) Control 2		Cytokine Panel 1 (human) Control 3	
		Recovery	CV	Recovery	CV	Recovery	CV	Recovery	CV
1	VEGF	80% - 120%	< 20%	90% - 130%	< 20%	90% - 130%	< 20%	90% - 130%	< 20%

Spot	Name	LLOQ			ULOQ		
		Value	Recovery	CV	Value	Recovery	CV
1	VEGF	7.7	80% - 120%	< 20%	562	80% - 120%	< 20%

**Stock Concentration of Calibrator:**

Spot	Name	Conc.	Units
1	VEGF	1820	pg/mL

Dilution to top of Curve: 1X

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
<b>Review/Approval</b>	Karen Hamilla	Quality		11 OCT 2024