

**CERTIFICATE OF ANALYSIS****Product Description:** V-PLEX® and V-PLEX Plus Proinflammatory Panel 1 (human) Kit**Lot Number:** K0082575**Expiration Date:** 31 MAR 2027**Catalog Number:** K151QR-Series; K151OD-Series; K15049-Series; K15052-Series; K15053-Series; K15346- Series; K15349-Series; K15345-Series; K15348-Series; K15056-Series**Lot-Specific Components:**

Description	Lot Number	Storage Temperature	Expiration Date
Anti-hu IFN- $\gamma$ Antibody	D0082286	2-8°C	31 DEC 2027
Anti-hu IL-10 Antibody	D0082251	2-8°C	30 SEP 2027
Anti-hu IL-12p70 Antibody	D0082235	2-8°C	30 SEP 2027
Anti-hu IL-13 Antibody	D0082187	2-8°C	30 JUN 2027
Anti-hu IL-1 $\beta$ Antibody	D0082176	2-8°C	30 JUN 2027
Anti-hu IL-2 Antibody	D0082312	2-8°C	31 JAN 2028
Anti-hu IL-4 Antibody	D0082287	2-8°C	31 DEC 2027
Anti-hu IL-6 Antibody	D0082249	2-8°C	30 SEP 2027
Anti-hu IL-8 Antibody	D0082274	2-8°C	30 NOV 2027
Anti-hu TNF- $\alpha$ Antibody	D0082288	2-8°C	31 DEC 2027
Proinflammatory Panel 1 (human) 96-Well 10-Spot SECTOR Plate	Z0048034	2-8°C	31 MAR 2027
Proinflammatory Panel 1 (human) Calibrator Blend	A0080411	2-8°C	31 OCT 2029
Diluent 2	M0020132	$\leq -10^{\circ}\text{C}$	31 MAR 2027
Diluent 3	M0030132	$\leq -10^{\circ}\text{C}$	31 MAR 2027

Additional components may be provided to execute the assay. Additional components are not lot-specific.

**Precision, Accuracy, and Sensitivity Specifications:**

Spot	Name	Calibrator		Proinflammatory Panel 1 (human) Control 1		Proinflammatory Panel 1 (human) Control 2		Proinflammatory Panel 1 (human) Control 3	
		Recovery	CV	Recovery	CV	Recovery	CV	Recovery	CV
1	IFN- $\gamma$	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
2	IL-1 $\beta$	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
3	IL-2	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
4	IL-4	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
5	IL-6	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
6	IL-8	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
7	IL-10	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
8	IL-12p70	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
9	IL-13	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
10	TNF- $\alpha$	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%



## CERTIFICATE OF ANALYSIS

Spot	Name	LLOQ			ULOQ		
		Value	Recovery	CV	Value	Recovery	CV
1	IFN- $\gamma$	1.76	80% - 120%	< 20%	938	80% - 120%	< 20%
2	IL-1 $\beta$	0.646	80% - 120%	< 20%	375	80% - 120%	< 20%
3	IL-2	0.89	80% - 120%	< 20%	938	80% - 120%	< 20%
4	IL-4	0.218	80% - 120%	< 20%	158	80% - 120%	< 20%
5	IL-6	0.633	80% - 120%	< 20%	488	80% - 120%	< 20%
6	IL-8	0.591	80% - 120%	< 20%	375	80% - 120%	< 20%
7	IL-10	0.298	80% - 120%	< 20%	233	80% - 120%	< 20%
8	IL-12p70	1.22	80% - 120%	< 20%	315	80% - 120%	< 20%
9	IL-13	4.21	80% - 120%	< 20%	353	80% - 120%	< 20%
10	TNF- $\alpha$	0.69	80% - 120%	< 20%	248	80% - 120%	< 20%

## Non-Specific Binding Specifications:

Spot	Name	NSB Spec
1	IFN- $\gamma$	< 1.0%
2	IL-1 $\beta$	< 1.0%
3	IL-2	< 1.0%
4	IL-4	< 1.0%
5	IL-6	< 1.0%
6	IL-8	< 1.0%
7	IL-10	< 1.0%
8	IL-12p70	< 1.0%
9	IL-13	< 1.0%
10	TNF- $\alpha$	< 1.0%

## Stock Concentration of Calibrator:

Spot	Name	Conc.	Units
1	IFN- $\gamma$	1410	pg/mL
2	IL-1 $\beta$	577	pg/mL
3	IL-2	1410	pg/mL
4	IL-4	220	pg/mL
5	IL-6	751	pg/mL
6	IL-8	603	pg/mL
7	IL-10	385	pg/mL
8	IL-12p70	507	pg/mL
9	IL-13	534	pg/mL
10	TNF- $\alpha$	369	pg/mL

Dilution to top of Curve: 1X

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Review/Approval	Karen Hamilla	Quality		28 JAN 2025