

**CERTIFICATE OF ANALYSIS****Product Description:** V-PLEX® and V-PLEX Plus Human GM-CSF Kit**Lot Number:** K0082641**Expiration Date:** 30 SEP 2026**Catalog Number:** K151RI-Series; K156RI-Series**Lot-Specific Components:**

Description	Lot Number	Storage Temperature	Expiration Date
Anti-hu GM-CSF Antibody	D0082095	2-8°C	31 DEC 2026
Cytokine Panel 1 (human) Calibrator Blend	A0080383	2-8°C	31 DEC 2028
Human GM-CSF 96-Well Small Spot SECTOR Plate	Z0048075	2-8°C	31 JUL 2027
Diluent 3	M0030136	≤ -10°C	31 MAY 2027
Diluent 43	M0430056	≤ -10°C	30 SEP 2026

Additional components may be provided to execute the assay. Additional components are not lot-specific.

**Precision, Accuracy, and Sensitivity Specifications:**

Spot	Name	Calibrator		Cytokine Panel 1 (human) Control 1		Cytokine Panel 1 (human) Control 2		Cytokine Panel 1 (human) Control 3	
		Recovery	CV	Recovery	CV	Recovery	CV	Recovery	CV
1	GM-CSF	80% - 120%	< 20%	90% - 130%	< 20%	90% - 130%	< 20%	90% - 130%	< 20%

Spot	Name	LLOQ			ULOQ		
		Value	Recovery	CV	Value	Recovery	CV
1	GM-CSF	0.84233	80% - 120%	< 20%	750	80% - 120%	< 20%

**Stock Concentration of Calibrator:**

Spot	Name	Conc.	Units
1	GM-CSF	1110	pg/mL

Dilution to top of Curve: 1X

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
<b>Review/Approval</b>	Michelle Likiam	Quality		25 FEB 2025