



CERTIFICATE OF ANALYSIS

Product Description: V-PLEX® and V-PLEX Plus Cytokine Panel 1 (human) Kit

Lot Number: K0082669

Expiration Date: 28 FEB 2027

Catalog Number: K15050-Series; K156RC-Series; K151RC-Series; K156RE-Series; K151RE-Series; K156RG-Series; K151RG-Series; K15057-Series

Lot-Specific Components:

Description	Lot Number	Storage Temperature	Expiration Date
Cytokine Panel 1 (human) 96-Well 10-Spot SECTOR Plate	Z0048095	2-8°C	31 AUG 2027
Cytokine Panel 1 (human) Calibrator Blend	A0080383	2-8°C	31 DEC 2028
Diluent 43	M0430058	≤ -10°C	28 FEB 2027
Diluent 3	M0030140	≤ -10°C	30 SEP 2027
Anti-hu GM-CSF Antibody	D0082279	2-8°C	31 DEC 2027
Anti-hu IL-1α Antibody	D0082280	2-8°C	31 JAN 2028
Anti-hu IL-5 Antibody	D0082281	2-8°C	30 NOV 2027
Anti-hu IL-7 Antibody	D0082346	2-8°C	29 FEB 2028
Anti-hu IL-12/23p40 Antibody	D0082282	2-8°C	30 NOV 2027
Anti-hu IL-15 Antibody	D0082283	2-8°C	30 NOV 2027
Anti-hu IL-16 Antibody	D0082349	2-8°C	29 FEB 2028
Anti-hu IL-17A Antibody	D0082350	2-8°C	29 FEB 2028
Anti-hu TNF-β Antibody	D0082351	2-8°C	29 FEB 2028
Anti-hu VEGF Antibody	D0082439	2-8°C	31 AUG 2028

Additional components may be provided to execute the assay. Additional components are not lot-specific.

Precision, Accuracy, and Sensitivity Specifications:

Spot	Name	Calibrator		Cytokine Panel 1 (human) Control 1		Cytokine Panel 1 (human) Control 2		Cytokine Panel 1 (human) Control 3	
		Recovery	CV	Recovery	CV	Recovery	CV	Recovery	CV
1	GM-CSF	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
2	IL-1α	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
3	IL-5	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
4	IL-7	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
5	IL-12/IL-23p40	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
6	IL-15	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
7	IL-16	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
8	IL-17A	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
9	TNF-β	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
10	VEGF-A	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%

Spot	Name	LLOQ			ULOQ		
		Value	Recovery	CV	Value	Recovery	CV
1	GM-CSF	0.842	80% - 120%	< 20%	750	80% - 120%	< 20%
2	IL-1α	2.85	80% - 120%	< 20%	278	80% - 120%	< 20%
3	IL-5	4.41	80% - 120%	< 20%	562	80% - 120%	< 20%
4	IL-7	0.851	80% - 120%	< 20%	563	80% - 120%	< 20%
5	IL-12/IL-23p40	1.32	80% - 120%	< 20%	2250	80% - 120%	< 20%
6	IL-15	0.774	80% - 120%	< 20%	525	80% - 120%	< 20%
7	IL-16	19.1	80% - 120%	< 20%	1870	80% - 120%	< 20%
8	IL-17A	3.19	80% - 120%	< 20%	3650	80% - 120%	< 20%
9	TNF-β	0.465	80% - 120%	< 20%	458	80% - 120%	< 20%
10	VEGF-A	7.70	80% - 120%	< 20%	562	80% - 120%	< 20%

**CERTIFICATE OF ANALYSIS****Non-Specific Binding Specifications:**

Spot	Name	GM-CSF	IL-1 α	IL-5	IL-7	IL-12/ IL-23p40	IL-15	IL-16	IL-17A	TNF- β	VEGF-A
1	GM-CSF	NA	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%
2	IL-1 α	< 1.0%	NA	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%
3	IL-5	< 1.0%	< 1.0%	NA	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%
4	IL-7	< 1.0%	< 1.0%	< 1.0%	NA	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%
5	IL-12/ IL-23p40	< 1.0%	< 1.0%	< 1.0%	< 1.0%	NA	< 1.0%	< 1.0%	< 3.0%	< 1.0%	< 1.0%
6	IL-15	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	NA	< 1.0%	< 1.0%	< 1.0%	< 1.0%
7	IL-16	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	NA	< 1.0%	< 1.0%	< 1.0%
8	IL-17A	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	NA	< 1.0%	< 1.0%
9	TNF- β	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	NA	< 1.0%
10	VEGF-A	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	< 1.0%	NA

Stock Concentration of Calibrator:

Spot	Name	Conc.	Units
1	GM-CSF	1110	pg/mL
2	IL-1 α	412	pg/mL
3	IL-5	826	pg/mL
4	IL-7	887	pg/mL
5	IL-12/IL-23p40	3460	pg/mL
6	IL-15	850	pg/mL
7	IL-16	2360	pg/mL
8	IL-17A	6490	pg/mL
9	TNF- β	721	pg/mL
10	VEGF-A	1820	pg/mL

Dilution to top of Curve: 1X

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Review/Approval	Karen Hamilla	Quality		16 MAY 2025