

**CERTIFICATE OF ANALYSIS****Product Description:** V-PLEX® and V-PLEX Plus Human IL-15 Kit**Lot Number:** K0082734**Expiration Date:** 28 FEB 2027**Catalog Number:** K151RD-Series; K156RD-Series**Lot-Specific Components:**

Description	Lot Number	Storage Temperature	Expiration Date
Anti-hu IL-15 Antibody	D0082181	2-8°C	30 JUN 2027
Cytokine Panel 1 (human) Calibrator Blend	A0080435	2-8°C	31 MAY 2030
Human IL-15 Plate	Z0048140	2-8°C	30 NOV 2027
Diluent 3	M0030140	≤ -10°C	30 SEP 2027
Diluent 43	M0430058	≤ -10°C	28 FEB 2027

Additional components may be provided to execute the assay. Additional components are not lot-specific.

Precision, Accuracy, and Sensitivity Specifications:

Spot	Name	Calibrator		Cytokine Panel 1 (human) Control 1		Cytokine Panel 1 (human) Control 2		Cytokine Panel 1 (human) Control 3	
		Recovery	CV	Recovery	CV	Recovery	CV	Recovery	CV
1	IL-15	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%


Spot	Name	LLOQ			ULOQ		
		Value	Recovery	CV	Value	Recovery	CV
1	IL-15	0.774	80% - 120%	< 20%	525	80% - 120%	< 20%

Stock Concentration of Calibrator:

Spot	Name	Conc.	Units
1	IL-15	889	pg/mL

Dilution to top of Curve: 1X

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Review/Approval	Michelle Likiam	Quality		21 AUG 2025