



**CERTIFICATE OF ANALYSIS**

**Product Description:** V-PLEX® and V-PLEX Plus Proinflammatory Panel 1 (human) Kit  
**Lot Number:** K0082817  
**Expiration Date:** 30 SEP 2027  
**Catalog Number:** K15049-Series, K15052-Series, K15053-Series, K15056-Series, K151QR-Series, K151OD-Series, K15345-Series, K15348-Series, K15346-Series, K15349-Series, K156AR-Series

**Lot-Specific Components:**

Description	Lot Number	Storage Temperature	Expiration Date
Anti-hu IFN-γ Antibody	D0082502	2-8°C	31 JAN 2029
Anti-hu IL-10 Antibody	D0082505	2-8°C	31 JAN 2029
Anti-hu IL-12p70 Antibody	D0082433	2-8°C	31 AUG 2028
Anti-hu IL-13 Antibody	D0082252	2-8°C	30 SEP 2027
Anti-hu IL-1β Antibody	D0082366	2-8°C	31 MAR 2028
Anti-hu IL-2 Antibody	D0082470	2-8°C	31 OCT 2028
Anti-hu IL-4 Antibody	D0082413	2-8°C	31 JUL 2028
Anti-hu IL-6 Antibody	D0082503	2-8°C	31 JAN 2029
Anti-hu IL-8 Antibody	D0082274	2-8°C	30 NOV 2027
Anti-hu TNF-α Antibody	D0082486	2-8°C	31 OCT 2028
Proinflammatory Panel 1 (hu) 96-Well 10-Spot SECTOR Plate	Z0048185	2-8°C	31 MAR 2028
Proinflammatory Panel 1 (human) Calibrator Blend	A0080434	2-8°C	31 MAY 2030
Diluent 2	M0020135	≤ -10°C	30 NOV 2027
Diluent 3	M0030141	≤ -10°C	30 NOV 2027

Additional components may be provided to execute the assay. Additional components are not lot-specific.

**Precision, Accuracy, and Sensitivity Specifications:**

Spot	Name	Calibrator		Proinflammatory Panel 1 (human) Control 1		Proinflammatory Panel 1 (human) Control 2		Proinflammatory Panel 1 (human) Control 3	
		Recovery	CV	Recovery	CV	Recovery	CV	Recovery	CV
1	IFN-γ	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
2	IL-1β	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
3	IL-2	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
4	IL-4	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
5	IL-6	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
6	IL-8	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
7	IL-10	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
8	IL-12p70	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
9	IL-13	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%
10	TNF-α	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%	80% - 120%	< 20%



### CERTIFICATE OF ANALYSIS

Spot	Name	LLOQ			ULOQ		
		Value	Recovery	CV	Value	Recovery	CV
1	IFN- $\gamma$	1.76	80% - 120%	< 20%	938	80% - 120%	< 20%
2	IL-1 $\beta$	0.646	80% - 120%	< 20%	375	80% - 120%	< 20%
3	IL-2	0.890	80% - 120%	< 20%	938	80% - 120%	< 20%
4	IL-4	0.218	80% - 120%	< 20%	158	80% - 120%	< 20%
5	IL-6	0.633	80% - 120%	< 20%	488	80% - 120%	< 20%
6	IL-8	0.591	80% - 120%	< 20%	375	80% - 120%	< 20%
7	IL-10	0.298	80% - 120%	< 20%	233	80% - 120%	< 20%
8	IL-12p70	1.22	80% - 120%	< 20%	315	80% - 120%	< 20%
9	IL-13	4.21	80% - 120%	< 20%	353	80% - 120%	< 20%
10	TNF- $\alpha$	0.690	80% - 120%	< 20%	248	80% - 120%	< 20%

#### Non-Specific Binding Specifications:

Spot	Name	Spec
1	IFN- $\gamma$	< 1.0%
2	IL-1 $\beta$	< 1.0%
3	IL-2	< 1.0%
4	IL-4	< 1.0%
5	IL-6	< 1.0%
6	IL-8	< 1.0%
7	IL-10	< 1.0%
8	IL-12p70	< 1.0%
9	IL-13	< 1.0%
10	TNF- $\alpha$	< 1.0%

#### Stock Concentration of Calibrator:

Spot	Name	Conc.	Units
1	IFN- $\gamma$	1170	pg/mL
2	IL-1 $\beta$	622	pg/mL
3	IL-2	1560	pg/mL
4	IL-4	276	pg/mL
5	IL-6	780	pg/mL
6	IL-8	555	pg/mL
7	IL-10	381	pg/mL
8	IL-12p70	494	pg/mL
9	IL-13	596	pg/mL
10	TNF- $\alpha$	346	pg/mL

Dilution to top of Curve: 1X

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Review/Approval	Karen Hamilla	Quality		20 NOV 2025