



Extension Memo

Page 1 of 1

Product Description: S-PLEX® Human Eotaxin-3 Kit
Catalog Number(s): K151Q3S-Series
Lot Number: K00S0184
New Expiration Date: 31 Jan 2026

Based on additional testing, the original date of expiration for this kit has been extended by 5 months.

The expiration dates for the following components have also been extended:

Description	Lot Number	Original Expiration Date	New Expiration Date
Human Eotaxin-3 Calibrator	A00S0028	31 Aug 2025	31 Jan 2026

Function	Name	Signature	Date
Quality	Karen Hamilla		05 May 2025

**CERTIFICATE OF ANALYSIS**

Product Description: S-PLEX[®] Human Eotaxin-3 Kit
Lot Number: K00S0184
Expiration Date: 31 AUG 2025
Catalog Number: K151Q3S-Series

Lot-Specific Components:

Description	Lot Number	Storage Temperature	Expiration Date
Biotin Human Eotaxin-3 Antibody	B00T0197	2-8°C	30 JUN 2028
Human Eotaxin-3 Calibrator	A00S0028	2-8°C	31 AUG 2025
TURBO-BOOST Human Eotaxin-3 Antibody	D00T0197	2-8°C	31 JUL 2028
Blocker S1	R0140013	≤ -10°C	31 JUL 2027
Diluent 57	M0570016	≤ -10°C	31 DEC 2026
S-PLEX Enhance E1	Y0550022	≤ -10°C	31 JUL 2027
S-PLEX Enhance E2	Y0560021	≤ -10°C	30 APR 2027
S-PLEX Coating Reagent C1	521329	≤ -70°C	31 DEC 2027
S-PLEX Detect D1 with TURBO-TAG [™] Reagent	Y0580030	≤ -70°C	30 APR 2027
S-PLEX Detect D2	Y0590023	≤ -70°C	31 JUL 2027
S-PLEX Enhance E3	Y0570024	≤ -70°C	30 JUN 2027

Additional components may be provided to execute the assay. Additional components are not lot-specific.

Precision, Accuracy, and Sensitivity Specifications:

Spot Number	Name	Calibrator	
		Recovery	CV
1	Eotaxin-3	80% - 120%	< 20%

Name	Range Tested	LLOQ				LLOD	
		Observed	Units	Recovery	CV	Observed	Units
Eotaxin-3	35.3 - 564	70.5	fg/mL	80% - 120%	< 20%	19.1	fg/mL

Stock Concentration of Calibrator:

Name	Item	Concentration	Units
Eotaxin-3	1X Calibrator	609000	fg/mL

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Review/Approval	Karen Hamilla	Quality		24 APR 2025