




## CERTIFICATE OF ANALYSIS

**Product Description:** MSD GOLD Read Buffer A  
**Catalog Number:** R92TG-Series  
**Lot Number:** Y0600072  
**Expiration Date:** 31 MAY 2029

**Statement:** MSD GOLD Read Buffer A is verified to ensure that consistent signals are obtained across lots. Details on the quality control testing of MSD GOLD Read Buffer A are provided in the product insert available at [www.mesoscale.com](http://www.mesoscale.com).

| Concentration of BTI    | Specification Metric               | Allowable Range  | Pass/Fail? |
|-------------------------|------------------------------------|------------------|------------|
| 2.00 nM BTI (0.10 pmol) | Test : Reference Signal Ratio      | 90 - 110 %       | Pass       |
|                         | Coefficient of Variation           | $\leq 10 \%$     | Pass       |
| 0 nM BTI (0 pmol)       | Test - Reference Signal Difference | $\pm 10$ counts  | Pass       |
|                         | Standard Deviation                 | $\leq 15$ counts | Pass       |

This product is intended for Research Use Only. Not for use in Diagnostic Procedures.

|                 | Name               | Function | Signature                                                                            | Date        |
|-----------------|--------------------|----------|--------------------------------------------------------------------------------------|-------------|
| Review/Approval | Carissa B. Hubbard | Quality  |  | 14 JUL 2026 |